

Sample Name: **Zprite LLR (Indigo-863) Primary**  
 Tested for: **Willamette Valley Alchemy**  
**Compliance Extract**

Laboratory ID: 20K0037-01

Matrix: Extracts and Concentrates

Sample Metrc ID: 1A4010300003909000013697

Lot # NA

Batch RFID: 1A4010300003909000013693

Batch Size: 1569 (g)

Process Date: 11/4/2020

License: 1000096CBB6

Date Sampled: 11/05/20 00:00

Date Accepted: 11/05/20




### Potency Analysis

Date Extracted: 11/10/20

Analysis Method: UNODC 5.4.8

Date Analyzed: 11/10/20

\* - ORELAP certified analyte

Cannabinoids	% weight	mg/g	LOQ (%)	Cannabinoids Profile														
<b>Total THC</b> ((THCA*0.877)+d9)	80.73	807.3	0.08	 <table border="1"> <tr><td>THC</td><td>80.73</td></tr> <tr><td>CBD</td><td>0.13</td></tr> <tr><td>CBG</td><td>1.19</td></tr> <tr><td>CBGA</td><td>0.14</td></tr> <tr><td>CBC</td><td>0.83</td></tr> <tr><td>THCV</td><td>0.38</td></tr> <tr><td>Total:</td><td>83.39</td></tr> </table>	THC	80.73	CBD	0.13	CBG	1.19	CBGA	0.14	CBC	0.83	THCV	0.38	Total:	83.39
THC	80.73																	
CBD	0.13																	
CBG	1.19																	
CBGA	0.14																	
CBC	0.83																	
THCV	0.38																	
Total:	83.39																	
<b>Total CBD</b> ((CBDA*0.877)+CBD)	0.13	1.3	0.08															
d9-THC (d9-Tetrahydrocannabinol)*	80.73	807.3	0.08															
d8-THC (d8-Tetrahydrocannabinol)*	< LOQ	< LOQ	0.10															
THCA (d9-Tetrahydrocannabinolic Acid)*	< LOQ	< LOQ	0.15															
CBD (Cannabidiol)*	0.13	1.3	0.08															
CBDA (Cannabidiolic Acid)*	< LOQ	< LOQ	0.15															
CBN (Cannabinol)*	< LOQ	< LOQ	0.08															
CBG (Cannabigerol)*	1.19	11.9	0.10															
CBGA (Cannabigerolic Acid)	0.14	1.4	0.10															
CBDV (Cannabidivarin)*	< LOQ	< LOQ	0.10															
CBDVA (Cannabidivarinic Acid)	< LOQ	< LOQ	0.10															
CBC (Cannabichromene)*	0.83	8.3	0.10															
THCV (Tetrahydrocannabivarin)	0.38	3.8	0.10															
<b>Total Cannabinoids</b>	<b>83.39</b>	<b>833.9</b>	<b>0.08</b>															

<LOQ - Results below the Limit of Quantitation - Compound not detected



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: **Zprite LLR (Indigo-863) Duplicate**  
 Tested for: **Willamette Valley Alchemy**  
**Compliance Extract**

Laboratory ID: 20K0037-02

Matrix: Extracts and Concentrates

Sample Metrc ID: 1A4010300003909000013697

Process Date: 11/4/2020

Lot # NA

License: 1000096CBB6

Batch RFID: 1A4010300003909000013693

Date Sampled: 11/05/20 00:00

Batch Size: 1569 (g)

Date Accepted: 11/05/20


### Potency Analysis

Date Extracted: 11/10/20

Analysis Method: UNODC 5.4.8

Date Analyzed: 11/10/20

\* - ORELAP certified analyte

Cannabinoids	% weight	mg/g	LOQ (%)	Cannabinoids Profile
<b>Total THC</b> ((THCA*0.877)+d9)	80.48	804.8	0.08	
<b>Total CBD</b> ((CBDA*0.877)+CBD)	0.15	1.5	0.08	
d9-THC (d9-Tetrahydrocannabinol)*	80.48	804.8	0.08	
d8-THC (d8-Tetrahydrocannabinol)*	< LOQ	< LOQ	0.10	
THCA (d9-Tetrahydrocannabinolic Acid)*	< LOQ	< LOQ	0.16	
CBD (Cannabidiol)*	0.15	1.5	0.08	
CBDA (Cannabidiolic Acid)*	< LOQ	< LOQ	0.16	
CBN (Cannabinol)*	0.08	0.8	0.08	
CBG (Cannabigerol)*	1.25	12.5	0.10	
CBGA (Cannabigerolic Acid)	< LOQ	< LOQ	0.10	
CBDV (Cannabidivarin)*	< LOQ	< LOQ	0.10	
CBDVA (Cannabidivarinic Acid)	< LOQ	< LOQ	0.10	
CBC (Cannabichromene)*	0.88	8.8	0.10	
THCV (Tetrahydrocannabivarin)	0.39	3.9	0.10	
<b>Total Cannabinoids</b>	<b>83.15</b>	<b>831.5</b>	<b>0.08</b>	

<LOQ - Results below the Limit of Quantitation - Compound not detected

Sample Name: **Zprite LLR (Indigo-863)**

Sample Metrc ID: **1A4010300003909000013697**

	Primary Result %	Duplicate Result %	Average %	% RPD	Pass/Fail (<15%RPD)
<b>Total THC</b> ((THCA*0.877)+d9)	80.73	80.48	80.61	0.31	PASS
<b>Total CBD</b> ((CBDA*0.877)+CBD)	0.13	0.15	0.14	NA	NA



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: <b>Zprite LLR (Indigo-863) Primary</b>	License: <b>100096CBB6</b>
Tested for: <b>Willamette Valley Alchemy Compliance Extract</b>	Date Sampled: <b>11/05/20 00:00</b>
Laboratory ID: <b>20K0037-01</b>	Date Accepted: <b>11/05/20</b>
Matrix: <b>Extracts and Concentrates</b>	Sample Metrc ID: <b>1A4010300003909000013697</b>
Lot # <b>NA</b>	Batch RFID: <b>1A4010300003909000013693</b>
	Batch Size: <b>1569 (g)</b>

### Terpene Analysis

Date Extracted: 11/10/20

Analysis Method: Terpenes by GC/FID

Date Analyzed: 11/11/20

Analyte	Result (%)	LOQ	Analyte	Result	LOQ
alpha Pinene	0.402	0.093	beta Myrcene	2.565	0.093
alpha Phellandrene	< LOQ	0.093	3-Carene	< LOQ	0.093
alpha Terpinene	< LOQ	0.093	Limonene	0.432	0.093
Terpinolene	1.284	0.093	Linalool	< LOQ	0.093
Fenchol	< LOQ	0.093	Borneol	< LOQ	0.093
Terpineol	0.146	0.093	Geraniol	< LOQ	0.093
alpha Humulene	0.204	0.093	beta Caryophyllene	0.589	0.093
(-)-Caryophyllene Oxide	< LOQ	0.093	(-)-alpha Bisabolol	0.217	0.093
Camphene	< LOQ	0.093	beta Pinene	0.203	0.093
Ocimene	< LOQ	0.093	Sabinene	< LOQ	0.093
Camphor	< LOQ	0.093	Isoborneol	< LOQ	0.093
Menthol	< LOQ	0.093	alpha Cedrene	< LOQ	0.093
Nerolidol	< LOQ	0.093	(+)-Pulegone	< LOQ	0.093
Eucalyptol	< LOQ	0.093	p-Cymene	< LOQ	0.093
(-)-Isopulegol	< LOQ	0.093	Geranyl Acetate	< LOQ	0.093
Guaiol	< LOQ	0.093	Valencene	< LOQ	0.093
Phytol	< LOQ	0.093	Citronellol	< LOQ	0.093
gamma Terpinene	< LOQ	0.093			
<b>Total Terpenes</b>				<b>6.041 %</b>	

<LOQ - Results below the Limit of Quantitation - Compound not detected

Terpene Analysis is not ORELAP Accredited.



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: **Zprite LLR (Indigo-863) Primary**

License: **1000096CBB6**

Tested for: **Willamette Valley Alchemy**

Date Sampled: **11/05/20 00:00**

**Compliance Extract**


Date Accepted: **11/05/20 16:39**

Laboratory ID: **20K0037-01**

Matrix: **Extracts and**

Client/Metrc ID: **1A4010300003909000013697**

**Terpene Profile**




Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: **Zprite LLR (Indigo-863) Primary** License: **100096CBB6**  
 Tested for: **Willamette Valley Alchemy** Date Sampled: **11/05/20 00:00**  
**Compliance Extract** Date Accepted: **11/05/20**

Laboratory ID: **20K0037-01** Sample Metrc ID: **1A4010300003909000013697**  
 Matrix: **Extracts and Concentrates** Batch RFID: **1A4010300003909000013693**  
 Lot # **NA** Batch Size: **1569 (g)**

## Pesticide Analysis in ppm

Date Extracted: 11/11/20 Analysis Method: AOAC 2007.01 & EN 15662  
 Date Analyzed: 11/12/20 Results above the action levels are highlighted in red #.

Analyte	Result	Action Level	LOQ	Analyte	Result	Action Level	LOQ
Abamectin	< LOQ	0.5	0.242	Acephate	< LOQ	0.4	0.194
Acequinocyl	< LOQ	2	0.969	Acetamiprid	< LOQ	0.2	0.097
Aldicarb	< LOQ	0.4	0.194	Azoxystrobin	< LOQ	0.2	0.097
Bifenazate	< LOQ	0.2	0.097	Bifenthrin	< LOQ	0.2	0.097
Boscalid	< LOQ	0.4	0.194	Carbaryl	< LOQ	0.2	0.097
Carbofuran	< LOQ	0.2	0.097	Chlorantraniliprole	< LOQ	0.2	0.097
Chlorfenapyr	< LOQ	1	0.484	Chlorpyrifos	< LOQ	0.2	0.097
Clofentezine	< LOQ	0.2	0.097	Cyfluthrin	< LOQ	1	0.484
Cypermethrin	< LOQ	1	0.484	Daminozide	< LOQ	1	0.484
DDVP (Dichlorvos)	< LOQ	1	0.484	Diazinon	< LOQ	0.2	0.097
Dimethoate	< LOQ	0.2	0.097	Ethoprophos	< LOQ	0.2	0.097
Etofenprox	< LOQ	0.4	0.194	Etoxazole	< LOQ	0.2	0.097
Fenoxycarb	< LOQ	0.2	0.097	Fenpyroximate	< LOQ	0.4	0.194
Fipronil	< LOQ	0.4	0.194	Fonicamid	< LOQ	1	0.484
Fludioxonil	< LOQ	0.4	0.194	Hexythiazox	< LOQ	1	0.484
Imazalil	< LOQ	0.2	0.097	Imidacloprid	< LOQ	0.4	0.194
Kresoxim-methyl	< LOQ	0.4	0.194	Malathion	< LOQ	0.2	0.097
Metalaxyl	< LOQ	0.2	0.097	Methiocarb	< LOQ	0.2	0.097
Methomyl	< LOQ	0.4	0.194	Methyl parathion	< LOQ	0.2	0.097
MGK-264	< LOQ	0.2	0.097	Myclobutanil	< LOQ	0.2	0.097
Naled	< LOQ	0.5	0.242	Oxamyl	< LOQ	1	0.484
Paclobutrazol	< LOQ	0.4	0.194	Permethrins (total)	< LOQ	0.2	0.097
Phosmet	< LOQ	0.2	0.097	Piperonyl butoxide	< LOQ	2	0.484
Prallethrin	< LOQ	0.2	0.097	Propiconazole	< LOQ	0.4	0.194
Propoxur	< LOQ	0.2	0.097	Pyrethrins (total)	< LOQ	1	0.484
Pyridaben	< LOQ	0.2	0.097	Spinosad	< LOQ	0.2	0.097
Spiromesifen	< LOQ	0.2	0.097	Spirotetramat	< LOQ	0.2	0.097
Spiroxamine	< LOQ	0.4	0.194	Tebuconazole	< LOQ	0.4	0.194
Thiacloprid	< LOQ	0.2	0.097	Thiamethoxam	< LOQ	0.2	0.097
Trifloxystrobin	< LOQ	0.2	0.097				

<LOQ - Results below the Limit of Quantitation - Compound not detected



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: **Zprite LLR (Indigo-863) Duplicate** License: **100096CBB6**  
 Tested for: **Willamette Valley Alchemy** Date Sampled: **11/05/20 00:00**  
**Compliance Extract** Date Accepted: **11/05/20**

Laboratory ID: **20K0037-02** Sample Metrc ID: **1A4010300003909000013697**  
 Matrix: **Extracts and Concentrates** Batch RFID: **1A4010300003909000013693**  
 Lot # **NA** Batch Size: **1569 (g)**

## Pesticide Analysis in ppm

Date Extracted: 11/11/20 Analysis Method: AOAC 2007.01 & EN 15662  
 Date Analyzed: 11/12/20 Results above the action levels are highlighted in red #.

Analyte	Result	Action Level	LOQ	Analyte	Result	Action Level	LOQ
Abamectin	< LOQ	0.5	0.240	Acephate	< LOQ	0.4	0.192
Acequinocyl	< LOQ	2	0.959	Acetamiprid	< LOQ	0.2	0.096
Aldicarb	< LOQ	0.4	0.192	Azoxystrobin	< LOQ	0.2	0.096
Bifenazate	< LOQ	0.2	0.096	Bifenthrin	< LOQ	0.2	0.096
Boscalid	< LOQ	0.4	0.192	Carbaryl	< LOQ	0.2	0.096
Carbofuran	< LOQ	0.2	0.096	Chlorantranilprole	< LOQ	0.2	0.096
Chlorfenapyr	< LOQ	1	0.480	Chlorpyrifos	< LOQ	0.2	0.096
Clofentezine	< LOQ	0.2	0.096	Cyfluthrin	< LOQ	1	0.480
Cypermethrin	< LOQ	1	0.480	Daminozide	< LOQ	1	0.480
DDVP (Dichlorvos)	< LOQ	1	0.480	Diazinon	< LOQ	0.2	0.096
Dimethoate	< LOQ	0.2	0.096	Ethoprophos	< LOQ	0.2	0.096
Etofenprox	< LOQ	0.4	0.192	Etoazole	< LOQ	0.2	0.096
Fenoxycarb	< LOQ	0.2	0.096	Fenpyroximate	< LOQ	0.4	0.192
Fipronil	< LOQ	0.4	0.192	Fonicamid	< LOQ	1	0.480
Fludioxonil	< LOQ	0.4	0.192	Hexythiazox	< LOQ	1	0.480
Imazalil	< LOQ	0.2	0.096	Imidacloprid	< LOQ	0.4	0.192
Kresoxim-methyl	< LOQ	0.4	0.192	Malathion	< LOQ	0.2	0.096
Metalaxyl	< LOQ	0.2	0.096	Methiocarb	< LOQ	0.2	0.096
Methomyl	< LOQ	0.4	0.192	Methyl parathion	< LOQ	0.2	0.096
MGK-264	< LOQ	0.2	0.096	Myclobutanil	< LOQ	0.2	0.096
Naled	< LOQ	0.5	0.240	Oxamyl	< LOQ	1	0.480
Paclobutrazol	< LOQ	0.4	0.192	Permethrins (total)	< LOQ	0.2	0.096
Phosmet	< LOQ	0.2	0.096	Piperonyl butoxide	< LOQ	2	0.480
Prallethrin	< LOQ	0.2	0.096	Propiconazole	< LOQ	0.4	0.192
Propoxur	< LOQ	0.2	0.096	Pyrethrins (total)	< LOQ	1	0.480
Pyridaben	< LOQ	0.2	0.096	Spinosad	< LOQ	0.2	0.096
Spiromesifen	< LOQ	0.2	0.096	Spirotetramat	< LOQ	0.2	0.096
Spiroxamine	< LOQ	0.4	0.192	Tebuconazole	< LOQ	0.4	0.192
Thiacloprid	< LOQ	0.2	0.096	Thiamethoxam	< LOQ	0.2	0.096
Trifloxystrobin	< LOQ	0.2	0.096				

<LOQ - Results below the Limit of Quantitation - Compound not detected



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: <b>Zprite LLR (Indigo-863) Primary</b>	License: <b>1000096CBB6</b>
Tested for: <b>Willamette Valley Alchemy Compliance Extract</b>	Date Sampled: <b>11/05/20 00:00</b>
	Date Accepted: <b>11/05/20</b>
Laboratory ID: <b>20K0037-01</b>	Sample Metric ID: <b>1A4010300003909000013697</b>
Matrix: <b>Extracts and Concentrates</b>	Batch RFID: <b>1A4010300003909000013693</b>
Lot # <b>NA</b>	Batch Size: <b>1569 (g)</b>

### Residual Solvents

Solvent	Results in ug/g	Action Level	LOQ	Date Extracted: 11/10/20
1,4-Dioxane	< LOQ	380	179	Date Analyzed: 11/11/20
2-Butanol	< LOQ	5000	2360	Analysis Method: USP 467
2-Ethoxyethanol	< LOQ	160	75.6	
2-Propanol (IPA)	< LOQ	5000	2360	
Acetone	< LOQ	5000	2360	
Acetonitrile	< LOQ	400	194	
Benzene	< LOQ	2	0.945	
Butanes	< LOQ	5000	2360	
Cyclohexane	< LOQ	3880	1830	
Dichloromethane (methylene chloride)	< LOQ	600	283	
Ethyl acetate	< LOQ	5000	2360	
Ethyl ether	< LOQ	5000	2360	
Ethylbenzene	< LOQ	2170	1020	
Ethylene glycol	< LOQ	620	293	
Ethylene oxide	< LOQ	50	23.6	
Heptane	< LOQ	5000	2360	
Hexanes	< LOQ	290	137	
Isopropyl acetate	< LOQ	5000	2360	
Isopropylbenzene (cumene)	< LOQ	70	33.1	
Methanol	< LOQ	3000	1420	
Pentanes	< LOQ	5000	2360	
Propane	< LOQ	5000	2360	
Tetrahydrofuran	< LOQ	720	340	
Toluene	< LOQ	890	420	
Xylenes	< LOQ	2170	1020	

<LOQ - Results below the Limit of Quantitation - Compound not detected  
 Results above the Action Level fail state testing requirements and will be highlighted **Red #**.



Brian Weigel  
 Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Sample Name: **Zprite LLR (Indigo-863) Duplicate**  
 Tested for: **Willamette Valley Alchemy**  
**Compliance Extract**

License: **100096CBB6**  
 Date Sampled: **11/05/20 00:00**  
 Date Accepted: **11/05/20**

Laboratory ID: **20K0037-02**

Sample Metric ID: **1A4010300003909000013697**

Matrix: **Extracts and Concentrates**

Batch RFID: **1A4010300003909000013693**

Lot # **NA**

Batch Size: **1569 (g)**

### Residual Solvents

Solvent	Results in ug/g	Action Level	LOQ
1,4-Dioxane	< LOQ	380	176
2-Butanol	< LOQ	5000	2310
2-Ethoxyethanol	< LOQ	160	74.1
2-Propanol (IPA)	< LOQ	5000	2310
Acetone	< LOQ	5000	2310
Acetonitrile	< LOQ	400	190
Benzene	< LOQ	2	0.926
Butanes	< LOQ	5000	2310
Cyclohexane	< LOQ	3880	1800
Dichloromethane (methylene chloride)	< LOQ	600	278
Ethyl acetate	< LOQ	5000	2310
Ethyl ether	< LOQ	5000	2310
Ethylbenzene	< LOQ	2170	1000
Ethylene glycol	< LOQ	620	287
Ethylene oxide	< LOQ	50	23.1
Heptane	< LOQ	5000	2310
Hexanes	< LOQ	290	134
Isopropyl acetate	< LOQ	5000	2310
Isopropylbenzene (cumene)	< LOQ	70	32.4
Methanol	< LOQ	3000	1390
Pentanes	< LOQ	5000	2310
Propane	< LOQ	5000	2310
Tetrahydrofuran	< LOQ	720	333
Toluene	< LOQ	890	412
Xylenes	< LOQ	2170	1000

Date Extracted: 11/10/20  
 Date Analyzed: 11/11/20  
 Analysis Method: USP 467

<LOQ - Results below the Limit of Quantitation - Compound not detected  
 Results above the Action Level fail state testing requirements and will be highlighted **Red #**.



Brian Weigel  
 Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.



**Case Narrative**

**Residual Solvent** - Isopropylbenzene was above normally accepted recovery criteria in the Matrix Spike and Matrix Spike Duplicate due to pinene coelution. Analyte was below the reporting limit in all client samples.

**Pesticides** - Multiple analytes recovered above the upper acceptance limit in the Blank Spike, Matrix Spike, and/or Matrix Spike Duplicate. Analytes were below the reporting limit in all client samples. See QC Report for details.

**Quality Control  
Potency**

**Batch: B202355 - Potency/Terpenes**

Blank(B202355-BLK1)		Extracted - 11/10/20 9:48 Analyzed - 11/10/20 18:49						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
d9-THC (d9-Tetrahydrocannabinol)	< LOQ	%						
d8-THC (d8-Tetrahydrocannabinol)	< LOQ	%						
THCA (d9-Tetrahydrocannabinolic Acid)	< LOQ	%						
CBD (Cannabidiol)	< LOQ	%						
CBDA (Cannabidiolic Acid)	< LOQ	%						
CBN (Cannabinol)	< LOQ	%						
CBG (Cannabigerol)	< LOQ	%						
CBGA (Cannabigerolic Acid)	< LOQ	%						
CBDV (Cannabidivarin)	< LOQ	%						
CBDVA (Cannabidivarinic Acid)	< LOQ	%						
CBC (Cannabichromene)	< LOQ	%						
THCV (Tetrahydrocannabivarin)	< LOQ	%						

Duplicate(B202355-DUP1)		Extracted - 11/10/20 9:48 Analyzed - 11/10/20 18:58						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
d9-THC (d9-Tetrahydrocannabinol)	77.79	%		79.55			2.24	20
d8-THC (d8-Tetrahydrocannabinol)	0.21	%		0.21			1.29	20
THCA (d9-Tetrahydrocannabinolic Acid)	< LOQ	%		< LOQ				20
CBD (Cannabidiol)	0.14	%		0.14			2.85	20
CBDA (Cannabidiolic Acid)	< LOQ	%		< LOQ				20
CBN (Cannabinol)	0.16	%		0.17			2.94	20
CBG (Cannabigerol)	0.81	%		0.83			3.16	20
CBGA (Cannabigerolic Acid)	< LOQ	%		< LOQ				20
CBDV (Cannabidivarin)	< LOQ	%		< LOQ				20
CBDVA (Cannabidivarinic Acid)	< LOQ	%		< LOQ				20
CBC (Cannabichromene)	0.82	%		0.81			1.33	20



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Potency (Continued)

**Batch: B202355 - Potency/Terpenes (Continued)**

Duplicate(B202355-DUP1)			Extracted - 11/10/20 9:48 Analyzed - 11/10/20 18:58					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
THCV (Tetrahydrocannabivarin)	0.46	%		0.46			0.286	20

LCS(B202355-BS1)			Extracted - 11/10/20 9:48 Analyzed - 11/10/20 18:41					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
d9-THC (d9-Tetrahydrocannabinol)	0.20	%	0.200		99.4	80-120		
CBD (Cannabidiol)	0.21	%	0.200		107	80-120		
CBDA (Cannabidiolic Acid)	0.18	%	0.200		87.7	80-120		
CBN (Cannabinol)	0.19	%	0.200		94.3	80-120		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Pesticide Analysis

**Batch: B202354 - Pesticide Prep**

Blank(B202354-BLK1)		Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:02						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Abamectin	< LOQ	ppm						
Acephate	< LOQ	ppm						
Acequinocyl	< LOQ	ppm						
Acetamiprid	< LOQ	ppm						
Aldicarb	< LOQ	ppm						
Azoxystrobin	< LOQ	ppm						
Bifenazate	< LOQ	ppm						
Bifenthrin	< LOQ	ppm						
Boscalid	< LOQ	ppm						
Carbaryl	< LOQ	ppm						
Carbofuran	< LOQ	ppm						
Chlorantraniliprole	< LOQ	ppm						
Chlorfenapyr	< LOQ	ppm						
Chlorpyrifos	< LOQ	ppm						
Clofentezine	< LOQ	ppm						
Cyfluthrin	< LOQ	ppm						
Cypermethrin	< LOQ	ppm						
Daminozide	< LOQ	ppm						
DDVP (Dichlorvos)	< LOQ	ppm						
Diazinon	< LOQ	ppm						
Dimethoate	< LOQ	ppm						
Ethoprophos	< LOQ	ppm						
Etofenprox	< LOQ	ppm						
Etoxazole	< LOQ	ppm						
Fenoxycarb	< LOQ	ppm						
Fenpyroximate	< LOQ	ppm						
Fipronil	< LOQ	ppm						
Fonicamid	< LOQ	ppm						
Fludioxonil	< LOQ	ppm						
Hexythiazox	< LOQ	ppm						
Imazalil	< LOQ	ppm						
Imidacloprid	< LOQ	ppm						
Kresoxim-methyl	< LOQ	ppm						
Malathion	< LOQ	ppm						
Metalaxyl	< LOQ	ppm						



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Pesticide Analysis (Continued)

**Batch: B202354 - Pesticide Prep (Continued)**

Blank(B202354-BLK1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:02					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Methiocarb	< LOQ	ppm						
Methomyl	< LOQ	ppm						
Methyl parathion	< LOQ	ppm						
MGK-264	< LOQ	ppm						
Myclobutanil	< LOQ	ppm						
Naled	< LOQ	ppm						
Oxamyl	< LOQ	ppm						
Paclbutrazol	< LOQ	ppm						
Permethrins (total)	< LOQ	ppm						
Phosmet	< LOQ	ppm						
Piperonyl butoxide	< LOQ	ppm						
Prallethrin	< LOQ	ppm						
Propiconazole	< LOQ	ppm						
Propoxur	< LOQ	ppm						
Pyrethrins (total)	< LOQ	ppm						
Pyridaben	< LOQ	ppm						
Spinosad	< LOQ	ppm						
Spiromesifen	< LOQ	ppm						
Spirotetramat	< LOQ	ppm						
Spiroxamine	< LOQ	ppm						
Tebuconazole	< LOQ	ppm						
Thiacloprid	< LOQ	ppm						
Thiamethoxam	< LOQ	ppm						
Trifloxystrobin	< LOQ	ppm						

LCS(B202354-BS1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:18					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Abamectin	1.17	ppm	0.980		119	15-180		
Acephate	1.09	ppm	1.00		109	51-141		
Acequinocyl	1.87	ppm	1.00		187	16.9-111		
Acetamiprid	1.13	ppm	1.00		113	50-150		
Aldicarb	1.11	ppm	1.00		111	49-146		
Azoxystrobin	1.08	ppm	1.00		108	52-136		
Bifenazate	1.15	ppm	1.00		115	41-133		
Bifenthrin	1.22	ppm	1.00		122	22-130		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Pesticide Analysis (Continued)

**Batch: B202354 - Pesticide Prep (Continued)**

LCS(B202354-BS1)		Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:18						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Boscalid	1.36	ppm	1.00		136	29-144		
Carbaryl	1.12	ppm	1.00		112	61-127		
Carbofuran	1.09	ppm	1.00		109	62-136		
Chlorantraniliprole	1.23	ppm	1.00		123	41-150		
Chlorfenapyr	1.27	ppm	1.00		127	40-160		
Chlorpyrifos	1.06	ppm	1.00		106	29-124		
Clofentezine	1.08	ppm	1.00		108	40-127		
Cyfluthrin	1.33	ppm	1.00		133	55-165		
Cypermethrin	1.37	ppm	1.00		137	21-144		
Daminozide	2.03	ppm	1.00		203	15-145		
DDVP (Dichlorvos)	1.05	ppm	1.00		105	55-150		
Diazinon	1.02	ppm	1.00		102	43-127		
Dimethoate	1.08	ppm	1.00		108	62-136		
Ethoprophos	1.07	ppm	1.00		107	45-142		
Etofenprox	1.15	ppm	1.00		115	24-113		
Etoxazole	1.15	ppm	1.00		115	34-121		
Fenoxycarb	1.23	ppm	1.00		123	22-150		
Fenpyroximate	1.50	ppm	1.00		150	34-144		
Fipronil	1.34	ppm	1.00		134	25-149		
Flonicamid	1.32	ppm	1.00		132	53-144		
Fludioxonil	0.89	ppm	1.00		89.4	29-132		
Hexythiazox	1.27	ppm	1.00		127	22-111		
Imazalil	1.03	ppm	1.00		103	48-125		
Imidacloprid	1.22	ppm	1.00		122	41-150		
Kresoxim-methyl	1.03	ppm	1.00		103	43-140		
Malathion	1.02	ppm	1.00		102	25-148		
Metalaxyl	1.12	ppm	1.00		112	50-136		
Methiocarb	1.07	ppm	1.00		107	56-132		
Methomyl	0.99	ppm	1.00		99.0	40-150		
Methyl parathion	1.15	ppm	1.00		115	35-160		
MGK-264	0.79	ppm	0.590		133	32-134		
Myclobutanil	1.47	ppm	1.00		147	43-141		
Naled	1.16	ppm	1.00		116	15-136		
Oxamyl	1.05	ppm	1.00		105	56-133		
Paclobutrazol	1.43	ppm	1.00		143	34-143		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Pesticide Analysis (Continued)

**Batch: B202354 - Pesticide Prep (Continued)**

LCS(B202354-BS1)		Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:18						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Permethrins (total)	1.29	ppm	1.00		129	31-113		
Phosmet	1.19	ppm	1.00		119	53-124		
Piperonyl butoxide	1.16	ppm	1.00		116	39-128		
Prallethrin	1.32	ppm	1.00		132	43-140		
Propiconazole	1.41	ppm	1.00		141	47-124		
Propoxur	1.09	ppm	1.00		109	63-135		
Pyrethrins (total)	0.77	ppm	0.580		133	19-144		
Pyridaben	1.22	ppm	1.00		122	31-122		
Spinosad	0.71	ppm	0.710		101	24-147		
Spiromesifen	1.08	ppm	1.00		108	49-133		
Spirotetramat	1.11	ppm	1.00		111	29-150		
Spiroxamine	1.26	ppm	1.00		126	15-122		
Tebuconazole	1.47	ppm	1.00		147	40-133		
Thiacloprid	1.15	ppm	1.00		115	60-143		
Thiamethoxam	1.15	ppm	1.00		115	42-146		
Trifloxystrobin	1.15	ppm	1.00		115	41-148		

Matrix Spike(B202354-MS1)		Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:34						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Abamectin	1.39	ppm	0.961	< LOQ	145	55-190		
Acephate	1.10	ppm	0.981	< LOQ	112	48-131		
Acequinocyl	0.61	ppm	0.981	< LOQ	62.2	15-119		
Acetamiprid	1.15	ppm	0.981	< LOQ	117	50-145		
Aldicarb	1.13	ppm	0.981	< LOQ	115	53-133		
Azoxystrobin	1.01	ppm	0.981	< LOQ	103	35-147		
Bifenazate	1.10	ppm	0.981	< LOQ	112	43-143		
Bifenthrin	0.10	ppm	0.981	< LOQ	10.5	16-107		
Boscalid	1.23	ppm	0.981	< LOQ	126	42-140		
Carbaryl	1.03	ppm	0.981	< LOQ	105	71-113		
Carbofuran	1.04	ppm	0.981	< LOQ	106	73-118		
Chlorantraniliprole	1.12	ppm	0.981	< LOQ	115	45-136		
Chlorfenapyr	0.91	ppm	0.981	< LOQ	92.5	40-190		
Chlorpyrifos	0.65	ppm	0.981	< LOQ	66.2	24-125		
Clofentezine	0.85	ppm	0.981	< LOQ	86.2	38-118		
Cyfluthrin	1.18	ppm	0.981	< LOQ	120	35-170		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Pesticide Analysis (Continued)

**Batch: B202354 - Pesticide Prep (Continued)**

Matrix Spike(B202354-MS1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:34					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Cypermethrin	1.00	ppm	0.981	< LOQ	102	38-150		
Daminozide	2.58	ppm	0.981	< LOQ	264	16-160		
DDVP (Dichlorvos)	1.30	ppm	0.981	< LOQ	132	64-124		
Diazinon	1.03	ppm	0.981	< LOQ	105	50-123		
Dimethoate	1.07	ppm	0.981	< LOQ	109	69-116		
Ethoprophos	0.97	ppm	0.981	< LOQ	99.1	39-146		
Etofenprox	0.43	ppm	0.981	< LOQ	43.4	31-117		
Etoxazole	0.96	ppm	0.981	< LOQ	97.9	35-136		
Fenoxycarb	1.17	ppm	0.981	< LOQ	120	23-150		
Fenpyroximate	1.62	ppm	0.981	< LOQ	165	30-143		
Fipronil	1.27	ppm	0.981	< LOQ	130	15-150		
Flonicamid	1.31	ppm	0.981	< LOQ	133	50-131		
Fludioxonil	0.90	ppm	0.981	< LOQ	91.4	44-150		
Hexythiazox	0.98	ppm	0.981	< LOQ	99.6	34-144		
Imazalil	1.06	ppm	0.981	< LOQ	108	54-124		
Imidacloprid	1.22	ppm	0.981	< LOQ	125	39-150		
Kresoxim-methyl	1.04	ppm	0.981	< LOQ	106	46-134		
Malathion	0.97	ppm	0.981	< LOQ	98.5	26-148		
Metalaxyl	1.08	ppm	0.981	< LOQ	110	60-127		
Methiocarb	1.05	ppm	0.981	< LOQ	107	50-131		
Methomyl	1.08	ppm	0.981	< LOQ	110	47-135		
Methyl parathion	1.02	ppm	0.981	< LOQ	104	33.5-156		
MGK-264	0.59	ppm	0.579	< LOQ	101	20-130		
Myclobutanil	1.36	ppm	0.981	< LOQ	139	43-134		
Naled	0.59	ppm	0.981	< LOQ	60.0	38-140		
Oxamyl	1.07	ppm	0.981	< LOQ	109	48-127		
Paclobutrazol	1.30	ppm	0.981	< LOQ	133	30-136		
Permethrins (total)	0.99	ppm	0.981	0.03	97.8	20-120		
Phosmet	1.09	ppm	0.981	< LOQ	112	51-134		
Piperonyl butoxide	0.81	ppm	0.981	< LOQ	82.7	36-134		
Prallethrin	1.28	ppm	0.981	< LOQ	130	23-149		
Propiconazole	1.50	ppm	0.981	< LOQ	153	45-133		
Propoxur	1.06	ppm	0.981	< LOQ	108	59-130		
Pyrethrins (total)	0.86	ppm	0.569	< LOQ	152	15-146		
Pyridaben	0.75	ppm	0.981	< LOQ	76.2	15-150		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Pesticide Analysis (Continued)

**Batch: B202354 - Pesticide Prep (Continued)**

Matrix Spike(B202354-MS1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20 22:34					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Spinosad	0.84	ppm	0.696	< LOQ	120	23-150		
Spiromesifen	0.91	ppm	0.981	< LOQ	92.7	27-127		
Spirotetramat	1.16	ppm	0.981	< LOQ	118	33-150		
Spiroxamine	1.28	ppm	0.981	< LOQ	130	54-134		
Tebuconazole	1.51	ppm	0.981	< LOQ	154	22-126		
Thiacloprid	1.09	ppm	0.981	< LOQ	112	53-138		
Thiamethoxam	1.12	ppm	0.981	< LOQ	115	40-134		
Trifloxystrobin	1.11	ppm	0.981	< LOQ	113	25-140		

Matrix Spike Dup(B202354-MSD1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Abamectin	1.54	ppm	0.980	< LOQ	157	55-190	8.20	40
Acephate	1.10	ppm	1.00	< LOQ	110	48-131	2.06	26
Acequinocyl	0.51	ppm	1.00	< LOQ	50.6	15-119	20.5	50
Acetamiprid	1.12	ppm	1.00	< LOQ	112	50-145	4.06	30
Aldicarb	1.11	ppm	1.00	< LOQ	111	53-133	3.80	30
Azoxystrobin	1.02	ppm	1.00	< LOQ	102	35-147	1.05	29
Bifenazate	1.11	ppm	1.00	< LOQ	111	43-143	0.355	30
Bifenthrin	0.11	ppm	1.00	< LOQ	10.9	16-107	3.69	29
Boscalid	1.28	ppm	1.00	< LOQ	128	42-140	2.03	30
Carbaryl	1.12	ppm	1.00	< LOQ	112	71-113	6.02	20
Carbofuran	1.11	ppm	1.00	< LOQ	111	73-118	4.30	20
Chlorantraniliprole	1.21	ppm	1.00	< LOQ	121	45-136	5.54	30
Chlorfenapyr	0.85	ppm	1.00	< LOQ	84.7	40-190	8.75	50
Chlorpyrifos	0.74	ppm	1.00	< LOQ	73.6	24-125	10.6	29
Clofentezine	0.99	ppm	1.00	< LOQ	99.0	38-118	13.8	26
Cyfluthrin	1.25	ppm	1.00	< LOQ	125	35-170	3.72	50
Cypermethrin	1.16	ppm	1.00	< LOQ	116	38-150	12.7	30
Daminozide	1.94	ppm	1.00	< LOQ	194	16-160	30.5	26
DDVP (Dichlorvos)	1.33	ppm	1.00	< LOQ	133	64-124	0.719	27
Diazinon	1.16	ppm	1.00	< LOQ	116	50-123	10.1	20
Dimethoate	1.08	ppm	1.00	< LOQ	108	69-116	0.973	20
Ethoprophos	1.05	ppm	1.00	< LOQ	105	39-146	5.56	30
Etofenprox	0.46	ppm	1.00	< LOQ	45.9	31-117	5.65	27
Etoxazole	0.89	ppm	1.00	< LOQ	89.0	35-136	9.57	30



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.



## Quality Control Pesticide Analysis (Continued)

Batch: B202354 - Pesticide Prep (Continued)

Matrix Spike Dup(B202354-MSD1)			Extracted - 11/11/20 12:45 Analyzed - 11/11/20					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
Fenoxycarb	1.24	ppm	1.00	< LOQ	124	23-150	3.38	40
Fenpyroximate	1.69	ppm	1.00	< LOQ	169	30-143	2.74	26
Fipronil	1.36	ppm	1.00	< LOQ	136	15-150	4.93	30
Flonicamid	1.36	ppm	1.00	< LOQ	136	50-131	2.29	26
Fludioxonil	0.94	ppm	1.00	< LOQ	94.1	44-150	2.90	30
Hexythiazox	0.90	ppm	1.00	< LOQ	90.0	34-144	10.1	28
Imazalil	1.09	ppm	1.00	< LOQ	109	54-124	0.704	24
Imidacloprid	1.22	ppm	1.00	< LOQ	122	39-150	2.63	30
Kresoxim-methyl	1.12	ppm	1.00	< LOQ	112	46-134	5.76	20
Malathion	1.01	ppm	1.00	< LOQ	101	26-148	2.81	50
Metalaxyl	1.13	ppm	1.00	< LOQ	113	60-127	2.37	30
Methiocarb	1.08	ppm	1.00	< LOQ	108	50-131	0.583	30
Methomyl	1.12	ppm	1.00	< LOQ	112	47-135	1.94	20
Methyl parathion	1.04	ppm	1.00	< LOQ	104	33.5-156	0.579	50
MGK-264	0.66	ppm	0.590	< LOQ	111	20-130	9.37	30
Myclobutanil	1.41	ppm	1.00	< LOQ	141	43-134	1.89	30
Naled	0.63	ppm	1.00	< LOQ	63.2	38-140	5.10	30
Oxamyl	1.08	ppm	1.00	< LOQ	108	48-127	1.06	28
Paclobutrazol	1.36	ppm	1.00	< LOQ	136	30-136	2.43	30
Permethrins (total)	1.07	ppm	1.00	0.03	104	20-120	6.21	28
Phosmet	1.18	ppm	1.00	< LOQ	118	51-134	5.71	30
Piperonyl butoxide	0.82	ppm	1.00	< LOQ	82.3	36-134	0.448	30
Prallethrin	1.35	ppm	1.00	< LOQ	135	23-149	3.97	30
Propiconazole	1.64	ppm	1.00	< LOQ	165	45-133	7.05	30
Propoxur	1.13	ppm	1.00	< LOQ	113	59-130	4.46	29
Pyrethrins (total)	0.93	ppm	0.580	< LOQ	160	15-146	5.29	28
Pyridaben	0.82	ppm	1.00	< LOQ	81.6	15-150	6.84	29
Spinosad	0.89	ppm	0.710	< LOQ	125	23-150	3.76	30
Spiromesifen	0.98	ppm	1.00	< LOQ	97.8	27-127	5.35	28
Spirotetramat	1.29	ppm	1.00	< LOQ	129	33-150	8.58	30
Spiroxamine	1.34	ppm	1.00	< LOQ	135	54-134	3.41	30
Tebuconazole	1.62	ppm	1.00	< LOQ	162	22-126	4.94	21
Thiacloprid	1.15	ppm	1.00	< LOQ	115	53-138	3.09	30
Thiamethoxam	1.14	ppm	1.00	< LOQ	114	40-134	0.731	28
Trifloxystrobin	1.18	ppm	1.00	< LOQ	118	25-140	3.93	30



Brian Weigel  
 Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Solvent Analysis

Batch: B202353 - Residual Solvent Prep

Blank(B202353-BLK1)			Extracted - 11/10/20 9:47 Analyzed - 11/10/20 22:49					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
1,4-Dioxane	< LOQ	ug/g						
2-Butanol	< LOQ	ug/g						
2-Ethoxyethanol	< LOQ	ug/g						
2-Propanol (IPA)	< LOQ	ug/g						
Acetone	< LOQ	ug/g						
Acetonitrile	< LOQ	ug/g						
Benzene	< LOQ	ug/g						
Butanes	< LOQ	ug/g						
Cyclohexane	< LOQ	ug/g						
Dichloromethane (methylene chloride)	< LOQ	ug/g						
Ethyl acetate	< LOQ	ug/g						
Ethyl ether	< LOQ	ug/g						
Ethylbenzene	< LOQ	ug/g						
Ethylene glycol	< LOQ	ug/g						
Ethylene oxide	< LOQ	ug/g						
Heptane	< LOQ	ug/g						
Hexanes	< LOQ	ug/g						
Isopropyl acetate	< LOQ	ug/g						
Isopropylbenzene (cumene)	< LOQ	ug/g						
Methanol	< LOQ	ug/g						
Pentanes	< LOQ	ug/g						
Propane	< LOQ	ug/g						
Tetrahydrofuran	< LOQ	ug/g						
Toluene	< LOQ	ug/g						
Xylenes	< LOQ	ug/g						

LCS(B202353-BS1)			Extracted - 11/10/20 9:47 Analyzed - 11/10/20 21:45					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
1,4-Dioxane	615	ug/g	570		108	70-130		
2,2-Dimethylbutane	513	ug/g	435		118	70-130		
2,2-Dimethylpropane (neopentane)	3880	ug/g	3120		124	60-140		
2-Butanol	3780	ug/g	3500		108	70-130		
2-Ethoxyethanol	228	ug/g	240		94.9	60-140		
2-Methylbutane (isopentane)	4290	ug/g	3500		123	70-130		
2-Methylpentane/2,3-Dimethylbutane	974	ug/g	870		112	70-130		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Solvent Analysis (Continued)

**Batch: B202353 - Residual Solvent Prep (Continued)**

LCS(B202353-BS1)		Extracted - 11/10/20 9:47 Analyzed - 11/10/20 21:45						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
2-Methylpropane (isobutane)	3790	ug/g	3120		121	60-140		
2-Propanol (IPA)	3970	ug/g	3500		113	70-130		
3-Methylpentane	478	ug/g	435		110	70-130		
Acetone	4080	ug/g	3500		117	70-130		
Acetonitrile	661	ug/g	615		107	70-130		
Benzene	2.64	ug/g	3.00		88.0	70-130		
Cyclohexane	6800	ug/g	5820		117	70-130		
Dichloromethane (methylene chloride)	920	ug/g	900		102	70-130		
Ethyl acetate	3940	ug/g	3500		113	70-130		
Ethyl ether	4150	ug/g	3500		119	70-130		
Ethylbenzene	3350	ug/g	3250		103	70-130		
Ethylene glycol	598	ug/g	930		64.3	60-140		
Ethylene oxide	415	ug/g	375		111	60-140		
Heptane	3970	ug/g	3500		113	70-130		
Isopropyl acetate	4020	ug/g	3500		115	70-130		
Isopropylbenzene (cumene)	108	ug/g	105		103	70-130		
m,p-Xylene	6870	ug/g	6510		106	60-140		
Methanol	2870	ug/g	2500		115	70-130		
n-Butane	3760	ug/g	3120		120	60-140		
n-Hexane	484	ug/g	435		111	70-130		
n-Pentane	4300	ug/g	3500		123	70-130		
Propane	1490	ug/g	1250		119	60-140		
Tetrahydrofuran	1250	ug/g	1080		116	70-130		
Toluene	1340	ug/g	1340		101	70-130		
o-Xylene	3330	ug/g	3250		102	70-130		

Matrix Spike(B202353-MS1)		Extracted - 11/10/20 9:47 Analyzed - 11/10/20 22:07						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
1,4-Dioxane	664	ug/g	643	< LOQ	103	70-130		
2,2-Dimethylbutane	568	ug/g	491	< LOQ	116	70-130		
2,2-Dimethylpropane (neopentane)	4360	ug/g	3530	< LOQ	124	60-140		
2-Butanol	4130	ug/g	3950	< LOQ	105	70-130		
2-Ethoxyethanol	241	ug/g	271	< LOQ	89.1	60-140		
2-Methylbutane (isopentane)	4770	ug/g	3950	< LOQ	121	70-130		
2-Methylpentane/2,3-Dimethylbutane	1080	ug/g	982	< LOQ	110	70-130		



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Solvent Analysis (Continued)

**Batch: B202353 - Residual Solvent Prep (Continued)**

Matrix Spike(B202353-MS1)			Extracted - 11/10/20 9:47 Analyzed - 11/10/20 22:07					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
2-Methylpropane (isobutane)	4310	ug/g	3530	< LOQ	122	60-140		
2-Propanol (IPA)	4330	ug/g	3950	< LOQ	110	70-130		
3-Methylpentane	530	ug/g	491	< LOQ	108	70-130		
Acetone	4480	ug/g	3950	< LOQ	113	70-130		
Acetonitrile	719	ug/g	694	< LOQ	104	70-130		
Benzene	2.95	ug/g	3.39	< LOQ	87.1	70-130		
Cyclohexane	7640	ug/g	6570	< LOQ	116	70-130		
Dichloromethane (methylene chloride)	1030	ug/g	1020	< LOQ	101	70-130		
Ethyl acetate	4300	ug/g	3950	< LOQ	109	70-130		
Ethyl ether	4600	ug/g	3950	< LOQ	117	70-130		
Ethylbenzene	3730	ug/g	3670	< LOQ	102	70-130		
Ethylene glycol	697	ug/g	1050	< LOQ	66.4	60-140		
Ethylene oxide	457	ug/g	423	< LOQ	108	60-140		
Heptane	4430	ug/g	3950	< LOQ	112	70-130		
Isopropyl acetate	4360	ug/g	3950	< LOQ	110	70-130		
Isopropylbenzene (cumene)	184	ug/g	118	< LOQ	155	70-130		
m,p-Xylene	7570	ug/g	7350	< LOQ	103	60-140		
Methanol	3190	ug/g	2820	< LOQ	113	70-130		
n-Butane	4210	ug/g	3530	< LOQ	119	60-140		
n-Hexane	539	ug/g	491	< LOQ	110	70-130		
n-Pentane	4790	ug/g	3950	< LOQ	121	70-130		
Propane	1650	ug/g	1410	< LOQ	117	60-140		
Tetrahydrofuran	1360	ug/g	1220	< LOQ	111	70-130		
Toluene	1510	ug/g	1510	< LOQ	99.8	70-130		
o-Xylene	3710	ug/g	3670	< LOQ	101	70-130		

Matrix Spike Dup(B202353-MSD1)			Extracted - 11/10/20 9:47 Analyzed - 11/10/20					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
1,4-Dioxane	662	ug/g	642	< LOQ	103	70-130	0.211	30
2,2-Dimethylbutane	574	ug/g	490	< LOQ	117	70-130	0.907	30
2,2-Dimethylpropane (neopentane)	4360	ug/g	3520	< LOQ	124	60-140	0.0534	30
2-Butanol	4140	ug/g	3940	< LOQ	105	70-130	0.0725	30
2-Ethoxyethanol	243	ug/g	270	< LOQ	89.8	60-140	0.621	30
2-Methylbutane (isopentane)	4830	ug/g	3940	< LOQ	123	70-130	1.36	30
2-Methylpentane/2,3-Dimethylbutane	1090	ug/g	980	< LOQ	112	70-130	1.49	30



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Solvent Analysis (Continued)

Batch: B202353 - Residual Solvent Prep (Continued)

Matrix Spike Dup(B202353-MSD1)			Extracted - 11/10/20 9:47 Analyzed - 11/10/20					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
2-Methylpropane (isobutane)	4300	ug/g	3520	< LOQ	122	60-140	0.138	30
2-Propanol (IPA)	4380	ug/g	3940	< LOQ	111	70-130	1.33	30
3-Methylpentane	538	ug/g	490	< LOQ	110	70-130	1.49	30
Acetone	4550	ug/g	3940	< LOQ	115	70-130	1.66	30
Acetonitrile	730	ug/g	693	< LOQ	105	70-130	1.52	30
Benzene	3.02	ug/g	3.38	< LOQ	89.4	70-130	2.48	30
Cyclohexane	7730	ug/g	6560	< LOQ	118	70-130	1.21	30
Dichloromethane (methylene chloride)	1040	ug/g	1010	< LOQ	103	70-130	1.48	30
Ethyl acetate	4340	ug/g	3940	< LOQ	110	70-130	1.06	30
Ethyl ether	4700	ug/g	3940	< LOQ	119	70-130	2.08	30
Ethylbenzene	3700	ug/g	3660	< LOQ	101	70-130	0.903	30
Ethylene glycol	697	ug/g	1050	< LOQ	66.5	60-140	0.00829	30
Ethylene oxide	464	ug/g	423	< LOQ	110	60-140	1.37	30
Heptane	4460	ug/g	3940	< LOQ	113	70-130	0.587	30
Isopropyl acetate	4370	ug/g	3940	< LOQ	111	70-130	0.439	30
Isopropylbenzene (cumene)	183	ug/g	118	< LOQ	154	70-130	0.605	30
m,p-Xylene	7600	ug/g	7340	< LOQ	104	60-140	0.367	30
Methanol	3240	ug/g	2820	< LOQ	115	70-130	1.54	30
n-Butane	4240	ug/g	3520	< LOQ	120	60-140	0.715	30
n-Hexane	547	ug/g	490	< LOQ	112	70-130	1.61	30
n-Pentane	4870	ug/g	3940	< LOQ	123	70-130	1.66	30
Propane	1650	ug/g	1410	< LOQ	117	60-140	0.395	30
Tetrahydrofuran	1360	ug/g	1220	< LOQ	112	70-130	0.364	30
Toluene	1490	ug/g	1510	< LOQ	99.0	70-130	0.985	30
o-Xylene	3720	ug/g	3660	< LOQ	102	70-130	0.305	30



Brian Weigel  
 Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control Terpene Analysis

Batch: B202356 - Potency/Terpenes

Blank(B202356-BLK1)			Extracted - 11/10/20 9:48 Analyzed - 11/10/20 22:03					
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
alpha Pinene	< LOQ	%						
beta Myrcene	< LOQ	%						
alpha Phellandrene	< LOQ	%						
3-Carene	< LOQ	%						
alpha Terpinene	< LOQ	%						
Limonene	< LOQ	%						
Terpinolene	< LOQ	%						
Linalool	< LOQ	%						
Fenchol	< LOQ	%						
Borneol	< LOQ	%						
Terpineol	< LOQ	%						
Geraniol	< LOQ	%						
alpha Humulene	< LOQ	%						
beta Caryophyllene	< LOQ	%						
(-)-Caryophyllene Oxide	< LOQ	%						
(-)-alpha Bisabolol	< LOQ	%						
Camphene	< LOQ	%						
beta Pinene	< LOQ	%						
Ocimene	< LOQ	%						
Sabinene	< LOQ	%						
Camphor	< LOQ	%						
Isoborneol	< LOQ	%						
Menthol	< LOQ	%						
alpha Cedrene	< LOQ	%						
Nerolidol	< LOQ	%						
(+)-Pulegone	< LOQ	%						
Eucalyptol	< LOQ	%						
p-Cymene	< LOQ	%						
(-)-Isopulegol	< LOQ	%						
Geranyl Acetate	< LOQ	%						
Guaiol	< LOQ	%						
Valencene	< LOQ	%						
Phytol	< LOQ	%						
Citronellol	< LOQ	%						
gamma Terpinene	< LOQ	%						



Brian Weigel  
 Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

## Quality Control

### Terpene Analysis (Continued)

**Batch: B202356 - Potency/Terpenes (Continued)**

Duplicate(B202356-DUP1)		Extracted - 11/10/20 9:48 Analyzed - 11/10/20 22:19						
Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit
alpha Pinene	< LOQ	%		< LOQ				30
beta Myrcene	0.539	%		0.583			7.89	30
alpha Phellandrene	< LOQ	%		< LOQ				30
3-Carene	< LOQ	%		< LOQ				30
alpha Terpinene	< LOQ	%		< LOQ				30
Limonene	0.936	%		1.006			7.21	30
Terpinolene	< LOQ	%		< LOQ				30
Linalool	0.230	%		0.240			4.38	30
Fenchol	0.218	%		0.232			6.05	30
Borneol	< LOQ	%		< LOQ				30
Terpineol	0.262	%		0.280			6.72	30
Geraniol	< LOQ	%		< LOQ				30
alpha Humulene	0.426	%		0.449			5.21	30
beta Caryophyllene	1.264	%		1.359			7.22	30
(-)-Caryophyllene Oxide	< LOQ	%		< LOQ				30
(-)-alpha Bisabolol	0.410	%		0.446			8.48	30
Camphene	< LOQ	%		< LOQ				30
beta Pinene	0.111	%		0.119			6.84	30
Ocimene	< LOQ	%		< LOQ				30
Sabinene	< LOQ	%		< LOQ				30
Camphor	< LOQ	%		< LOQ				30
Isoborneol	< LOQ	%		< LOQ				30
Menthol	< LOQ	%		< LOQ				30
alpha Cedrene	< LOQ	%		< LOQ				30
Nerolidol	< LOQ	%		< LOQ				30
(+)-Pulegone	< LOQ	%		< LOQ				30
Eucalyptol	< LOQ	%		< LOQ				30
p-Cymene	< LOQ	%		< LOQ				30
(-)-Isopulegol	< LOQ	%		< LOQ				30
Geranyl Acetate	< LOQ	%		< LOQ				30
Guaiol	< LOQ	%		< LOQ				30
Valencene	0.168	%		0.175			4.39	30
Phytol	< LOQ	%		< LOQ				30
Citronellol	< LOQ	%		< LOQ				30
gamma Terpinene	< LOQ	%		< LOQ				30



Brian Weigel  
Lab Director

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of SC Laboratories. Samples tested in accordance with Oregon Administrative Rules, TNI 2009 Standard and SC Laboratories quality assurance plan unless otherwise noted.

Client: Willamette Valley Alchemy Client License: 100096CBBC Date Sampled: 11/5/2020 Thermometer ID: T014  
 Address Where Sampled: 870 W 2nd Ave unit: D Eugene, OR 97402 Requestor: Paul Sherman Event ID: 20K0037 Balance ID: SAMP\_BAL\_02  
 Sampling SOP & Rev. #: SC-OR-SAMP-003 Sampler: Scott Forster Transporter: Scott Forster Hygrometer ID: Anemometer\_02  
 Lab ORELAP ID: 4133  
 Lab OLCG ID: 1004748743D

Sampler Signature



Weight used (g)	Weight Set ID	Acceptance Criteria	Initial Measured	Initial P/F	Final Measured	Final P/F
0.5	SAMP_CAL_02	±2.5%	0.5	P	0.5	P
200		±2.5%	199.96		199.96	

Container Type	METRC Harvest/Processing Lot ID #:				Product Type	Client Sample Name	Product Date	Batch Size (g)
Mason Jars	1A40103000390900013693				Concentrate	Zprite LLR (Indigo-863)	11/4/2020	1569
METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	4	vial	4	8	0.4375	Zprite LLR (Indigo-863) Primary
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-01	Zprite LLR (Indigo-863)-1		A2	0	0.44	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-1		A3	0.44	0.88	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-1		A3	0.88	1.32	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-1		A3	1.32	1.76	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-2		B2	1.76	2.2	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-3		C1	2.2	2.64	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-3		C4	2.64	3.08	0.44	13697	
20K0037-01	Zprite LLR (Indigo-863)-4		D1	3.08	3.57	0.49	13697	
<b>Totals:</b>								
8			8		Total Primary Mass = 3.57		Primary + Duplicate Mass = 7.16 g	

Observations and Abnormalities:	Batch #	Equipment	Cont. Types/Sizes	Uniform	Plant Colors	Shape and Size	Sampling Plan ID & Rev. Date

METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	4	vial	4	8	0.4375	Zprite LLR (Indigo-863) Duplicate
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-02	Zprite LLR (Indigo-863)-1		A1	0	0.44	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-1		A1	0.44	0.88	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-1		A2	0.88	1.32	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-1		A3	1.32	1.76	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-2		B3	1.76	2.2	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-2		B4	2.2	2.64	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-3		C3	2.64	3.08	0.44	13697	
20K0037-02	Zprite LLR (Indigo-863)-4		D2	3.08	3.59	0.51	13697	



Totals:		8		8		Total Duplicate Mass = 3.59	
Observations and Abnormalities:		Batch #	Equipment	Cont. Types/Sizes	Uniform	Plant Colors	Shape and Size
							Sampling Plan ID & Rev. Date

Container Type	METRC Harvest/Processing Lot ID #:				Product Type	Client Sample Name	Product Date	Batch Size (g)
Mason Jars	1A4010300003909000013694				Concentrate	Mac CRJ (Siren-866)	11/4/2020	794
METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	2	vial	4	6	0.58333333	Mac CRJ (Siren-866) Primary
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-03	Mac CRJ (Siren-866)-1		A2	0	0.59	0.59	13698	
20K0037-03	Mac CRJ (Siren-866)-2		B3	0.59	1.18	0.59	13698	
20K0037-03	Mac CRJ (Siren-866)-2		B4	1.18	1.77	0.59	13698	
20K0037-03	Mac CRJ (Siren-866)-2		B4	1.77	2.36	0.59	13698	
20K0037-03	Mac CRJ (Siren-866)-2		B4	2.36	2.95	0.59	13698	
20K0037-03	Mac CRJ (Siren-866)-2		B4	2.95	3.57	0.62	13698	
Totals:		6		6		Total Primary Mass = 3.57		Primary + Duplicate Mass = 7.12 g

Observations and Abnormalities:		Batch #	Equipment	Cont. Types/Sizes	Uniform	Plant Colors	Shape and Size	Sampling Plan ID & Rev. Date

METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	2	vial	4	6	0.58333333	Mac CRJ (Siren-866) Duplicate
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-04	Mac CRJ (Siren-866)-1		A2	0	0.59	0.59	13698	
20K0037-04	Mac CRJ (Siren-866)-1		A3	0.59	1.18	0.59	13698	
20K0037-04	Mac CRJ (Siren-866)-2		B1	1.18	1.77	0.59	13698	
20K0037-04	Mac CRJ (Siren-866)-2		B2	1.77	2.36	0.59	13698	
20K0037-04	Mac CRJ (Siren-866)-2		B2	2.36	2.95	0.59	13698	
20K0037-04	Mac CRJ (Siren-866)-2		B2	2.95	3.55	0.6	13698	

<b>Totals:</b>	6	6				Total Duplicate Mass = 3.55		Primary + Duplicate Mass = 7.12 g
<b>Observations and Abnormalities:</b>	<b>Batch #</b>	<b>Equipment</b>	<b>Cont. Types/Sizes</b>	<b>Uniform</b>	<b>Plant Colors</b>	<b>Shape and Size</b>	<b>Sampling Plan ID &amp; Rev. Date</b>	

Container Type	METRC Harvest/Processing Lot ID #:				Product Type	Client Sample Name	Product Date	Batch Size (g)
Mason Jars	1A4010300003909000013695				Concentrate	Unicorn Poop Badder (Focus-869)	11/4/2020	338
METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	1	vial	4	4	0.875	Unicorn Poop Badder (Focus-869) Primary
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-05	Unicorn Poop Badder (Focus-869)-1		A1	0	0.88	0.88	13699	
20K0037-05	Unicorn Poop Badder (Focus-869)-1		A3	0.88	1.76	0.88	13699	
20K0037-05	Unicorn Poop Badder (Focus-869)-1		A4	1.76	2.64	0.88	13699	
20K0037-05	Unicorn Poop Badder (Focus-869)-1		A4	2.64	3.54	0.9	13699	
<b>Totals:</b>	4	4				Total Primary Mass = 3.54		Primary + Duplicate Mass = 7.06 g
<b>Observations and Abnormalities:</b>	<b>Batch #</b>	<b>Equipment</b>	<b>Cont. Types/Sizes</b>	<b>Uniform</b>	<b>Plant Colors</b>	<b>Shape and Size</b>	<b>Sampling Plan ID &amp; Rev. Date</b>	
METRC Batch ID	Product Temp (°C)	Humidity (%)	# of Containers	Sampling Media	# Zones	# of Inc.	1° Sample (g)	Sample Name
	17.3	47.1	1	vial	4	4	0.875	Unicorn Poop Badder (Focus-869) Duplicate
Lab Sample ID	Container ID		Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-06	Unicorn Poop Badder (Focus-869)-1		A1	0	0.88	0.88	13699	
20K0037-06	Unicorn Poop Badder (Focus-869)-1		A2	0.88	1.76	0.88	13699	
20K0037-06	Unicorn Poop Badder (Focus-869)-1		A3	1.76	2.64	0.88	13699	
20K0037-06	Unicorn Poop Badder (Focus-869)-1		A4	2.64	3.52	0.88	13699	



Lab Sample ID	Container ID	Increment Zone	Sampling Media Wt. (g)	Wt. Inc+Media (g)	Increment Weight (g)	Sample METRC ID#	
20K0037-08	Donkey Butter Shatter (Tokyo-871)-1	A2	0	0.88	0.88	13700	
20K0037-08	Donkey Butter Shatter (Tokyo-871)-1	A2	0.88	1.76	0.88	13700	
20K0037-08	Donkey Butter Shatter (Tokyo-871)-1	A3	1.76	2.64	0.88	13700	
20K0037-08	Donkey Butter Shatter (Tokyo-871)-1	A3	2.64	3.55	0.91	13700	
<b>Totals:</b>		4	4	Total Duplicate Mass = 3.55		Primary + Duplicate Mass = 7.11 g	
Observations and Abnormalities:	Batch #	Equipment	Cont. Types/Sizes	Uniform	Plant Colors	Shape and Size	Sampling Plan ID & Rev. Date



**OREGON LIQUOR CONTROL COMMISSION  
CANNABIS TRANSPORTATION MANIFEST**



20K0037

All sales transactions are to be completed prior to transportation of any CANNABIS. The receiving entity may reject product delivered, but amount delivered must be limited to amount agreed upon in prior sales transaction.

<b>Manifest No.</b>	<b>0002793188</b>	<b>Date Created</b>	11/5/2020 2:21 PM
<b>Originating Entity</b>	Willamette Valley Alchemy	<b>For OLCC Use Only</b>	
<b>Originating License Number</b>	030-1000096CBB6		
<b>Address of Originating Entity</b>	870 W. 2nd Ave Unit: D Eugene, OR 97402 County: Lane		
<b>Phone No. of Originating Entity</b>	541255.9170		
<b>Contact Phone No. for Inquiries: 8018824601</b>			
<b>1. Destination</b>	SC Laboratories	<b>Destination Phone No.</b>	503-272-8838
<b>Destination License Number</b>	010-1004748743D	<b>Date and Approx. Time of Departure</b>	11/5/2020 2:30 PM
<b>Address of Destination</b>	15865 SW 74th Avenue Ste 110 Tigard, OR 97224 County: Washington	<b>Date and Approx. Time of Arrival</b>	11/5/2020 5:30 PM
		<b>Date/Time Received</b>	11-5-2020 16:39
		<b>Notes:</b> details for extenuating circumstances (e.g., road closure, flat tire, etc.)	
<b>Route to be Traveled</b> Get on I-105 E from Monroe St and W 5th Ave Follow I-5 N to Lower Boones Ferry Rd in Tualatin. Take exit 290 from I-5 N Take SW Durham Rd to SW 74th Ave in Tigard			
<b>Name of Person Transporting</b>	Scott Forster	<b>Handler Permit No. of Driver</b>	SC Sampler
<b>State Driver's License No.</b>	A625521	<b>Signature of Person Transporting</b>	
<b>Make, Model, License Plate No.</b>	Nissan Kicks 249 MGD		
<b>1. Package I Shipped</b>	<b>Production Batch No.</b>	<b>Item Name</b>	<b>Quantity</b>
1A4010300003909000013700 Lab Test: SubmittedForTesting		Donkey Butter Shatter (Tokyo-871) (Extracts)	Shp: 7.1100 g
<b>Item Details</b>			
<b>Source Harvest(s)</b>	(multi-harvest)		
<b>Source Package(s)</b>	1A4010300003909000013696		
<b>2. Package I Shipped</b>	<b>Production Batch No.</b>	<b>Item Name</b>	<b>Quantity</b>
1A4010300003909000013699 Lab Test: SubmittedForTesting		Unicorn Poop Badder (Focus-869) (Extracts)	Shp: 7.0600 g
<b>Item Details</b>			
<b>Source Harvest(s)</b>	Unicorn Poop F 8.13.20		
<b>Source Package(s)</b>	1A4010300003909000013695		
<b>3. Package I Shipped</b>	<b>Production Batch No.</b>	<b>Item Name</b>	<b>Quantity</b>
1A4010300003909000013697 Lab Test: SubmittedForTesting		Zprite LLR (Indigo-863) (Extracts)	Shp: 7.1600 g
<b>Item Details</b>			
<b>Source Harvest(s)</b>	(multi-harvest)		
<b>Source Package(s)</b>	1A4010300003909000013693		
<b>4. Package I Shipped</b>	<b>Production Batch No.</b>	<b>Item Name</b>	<b>Quantity</b>
1A4010300003909000013698 Lab Test: SubmittedForTesting		Mac CRJ (Siren-866) (Extracts)	Shp: 7.1200 g
<b>Item Details</b>			
<b>Source Harvest(s)</b>	(multi-harvest)		
<b>Source Package(s)</b>	1A4010300003909000013694		



**OREGON LIQUOR CONTROL COMMISSION  
CANNABIS TRANSPORTATION MANIFEST**



20K0037

All sales transactions are to be completed prior to transportation of any CANNABIS. The receiving entity may reject product delivered, but amount delivered must be limited to amount agreed upon in prior sales transaction.

<b>Manifest No.</b>	<b>0002793188</b>	<b>Date Created</b>	11/5/2020 2:21 PM
PRODUCT REJECTION (if only a portion of shipment is rejected, circle that portion above)			
<b>Name of Person Receiving or Rejecting Product</b>	Leonel Penza		
I confirm that the contents of this shipment match weight records entered above, and I agree to take custody of those portions of this shipment <i>not</i> circled above. Those portions circled were returned to the individual delivering this shipment.			
<b>Signature</b>	Leonel Penza	<b>Date</b>	11-5-2020
<b>Signature of individual taking receipt of rejected portion of this shipment</b>			



<b>CHAIN OF CUSTODY</b>		<b>Client</b>	<b>COC #</b>	<b>Sample Type Legend</b>
SC Laboratories Oregon LLC 15865 SW 74th Avenue, Ste 110 Tigard OR, 97224 (503) 272-8830 ORELAP ID #4133 www.sclabs.com		Williamette Valley Alchemy 870 W 2nd Ave unit: D Eugene, OR 97402	1/1 20K0037	U - Usable Marijuana C - Concentrate P - Product O - Other
<b>Address Where Sampled</b>	<b>Date Sampled</b>	<b>OLCC License #</b>	<b>Received By</b>	<b>Received Date</b>
15865 SW 74th Avenue, Ste 110 Tigard OR, 97224 (503) 272-8830 ORELAP ID #4133 www.sclabs.com	11/5/2020	100096CBBC	LJP	11-5-2020
<b>OLCC License Type</b>	<b>Processor</b>	<b>Name of Sampler</b>	<b>Courier</b>	<b>Place where Sampled</b>
OLCC License Type	100096CBBC	Scott Forster	Scott Forster	870 W 2nd Ave unit: D Eugene, OR 97402
<b>Email</b>	<b>Phone</b>	<b>Transfer Manifest #</b>	<b>Water Activity</b>	<b>Moisture Content</b>
stvalleyvalchemistry@gmail.com	541-255-9170	2793188		
<b>Sampler OLCC License #</b>	<b>METRC Label</b>	<b>Unique Batch Number</b>	<b>Sample Type</b>	<b>Total Sample Mass</b>
010-1004748743D	13697	Zprite LLR (Indigo-863)	C	3.57
	13697	Zprite LLR (Indigo-863)	C	3.59
	13698	Mac CRJ (Siren-866)	C	3.57
	13698	Mac CRJ (Siren-866)	C	3.55
	13699	Unicorn Poop Badder (Focus-869)	C	3.54
	13699	Unicorn Poop Badder (Focus-869)	C	3.52
	13700	Donkey Butter Shatter (Tokyo-871)	C	3.56
	13700	Donkey Butter Shatter (Tokyo-871)	C	3.55

Sample Name	Time	METRC Label	Unique Batch Number	SC Labs LIMS ID	Sample Type	Total Sample Mass	# of Increments	TESTS REQUESTED						Sample Specific Notes
								Potency	Water Activity	Moisture Content	Pesticide	Residual Solvent	Terpene	
Zprite LLR (Indigo-863) Primary		13697	Zprite LLR (Indigo-863)	20K0037-01	C	3.57	8	X	X	X	X	X	X	
Zprite LLR (Indigo-863) Duplicate		13697	Zprite LLR (Indigo-863)	20K0037-02	C	3.59	8	X	X	X	X	X	X	
Mac CRJ (Siren-866) Primary		13698	Mac CRJ (Siren-866)	20K0037-03	C	3.57	6	X	X	X	X	X	X	
Mac CRJ (Siren-866) Duplicate		13698	Mac CRJ (Siren-866)	20K0037-04	C	3.55	6	X	X	X	X	X	X	
Unicorn Poop Badder (Focus-869) Primary		13699	Unicorn Poop Badder (Focus-869)	20K0037-05	C	3.54	4	X	X	X	X	X	X	
Unicorn Poop Badder (Focus-869) Duplicate		13699	Unicorn Poop Badder (Focus-869)	20K0037-06	C	3.52	4	X	X	X	X	X	X	
Donkey Butter Shatter (Tokyo-871) Primary		13700	Donkey Butter Shatter (Tokyo-871)	20K0037-07	C	3.56	4	X	X	X	X	X	X	
Donkey Butter Shatter (Tokyo-871) Duplicate		13700	Donkey Butter Shatter (Tokyo-871)	20K0037-08	C	3.55	4	X	X	X	X	X	X	

**Notes/Special Considerations:** Opt OUT of Sample Duplicate Yes  No

Samples Relinquished	Samples Received
Print Name: Austin C    Date: 11/5/20 Representative of: W.V.A. Signature: <i>Austin C</i> Time: 230	Print Name: Scott F    Date: 11/5/20 Representative of: SC Labs Signature: <i>Scott F</i> Time: 230